

October 03, 2018

Rob King
Hampton Bays Water District
P.O. Box 1013
Hampton Bays, NY 11946

RE: Project: FE/MN 10/1
Pace Project No.: 7066406

Dear Rob King:

Enclosed are the analytical results for sample(s) received by the laboratory on October 01, 2018. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Stu Murrell
stu.murrell@pacelabs.com
(631)694-3040
Project Manager

Enclosures

cc: Warren Booth, Hampton Bays Water District
John Collins, H2M Group
Stella Michaels, Hampton Bays Water District
Paul Ponturo, H2M Group



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: FE/MN 10/1

Pace Project No.: 7066406

Long Island Certification IDs

575 Broad Hollow Rd, Melville, NY 11747

New York Certification #: 10478 Primary Accrediting Body

New Jersey Certification #: NY158

Pennsylvania Certification #: 68-00350

Connecticut Certification #: PH-0435

Maryland Certification #: 208

Rhode Island Certification #: LAO00340

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: FE/MN 10/1

Pace Project No.: 7066406

Lab ID	Sample ID	Matrix	Date Collected	Date Received
7066406001	S-108065	Drinking Water	10/01/18 10:31	10/01/18 15:45
7066406002	S-108066	Drinking Water	10/01/18 10:33	10/01/18 15:45
7066406003	COMB. ENTRY POINT	Drinking Water	10/01/18 10:33	10/01/18 15:45

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: FE/MN 10/1

Pace Project No.: 7066406

Lab ID	Sample ID	Method	Analysts	Analytes Reported
7066406001	S-108065	EPA 200.7	AKS	2
7066406002	S-108066	EPA 200.7	AKS	2
7066406003	COMB. ENTRY POINT	EPA 200.7	AKS	2

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: FE/MN 10/1

Pace Project No.: 7066406

Sample: S-108065		Lab ID: 7066406001		Collected: 10/01/18 10:31	Received: 10/01/18 15:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.7 MET ICP, Drinking Water		Analytical Method: EPA 200.7							
Iron	0.94	mg/L	0.020		1		10/02/18 11:52	7439-89-6	
Manganese	0.15	mg/L	0.010		1		10/02/18 11:52	7439-96-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: FE/MN 10/1

Pace Project No.: 7066406

Sample: S-108066		Lab ID: 7066406002		Collected: 10/01/18 10:33	Received: 10/01/18 15:45	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.7 MET ICP, Drinking Water		Analytical Method: EPA 200.7								
Iron	0.49	mg/L	0.020		1		10/02/18 11:56	7439-89-6		
Manganese	0.061	mg/L	0.010		1		10/02/18 11:56	7439-96-5		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: FE/MN 10/1

Pace Project No.: 7066406

Sample: COMB. ENTRY POINT		Lab ID: 7066406003		Collected: 10/01/18 10:33	Received: 10/01/18 15:45	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.7 MET ICP, Drinking Water		Analytical Method: EPA 200.7								
Iron	0.73	mg/L	0.020		1		10/02/18 12:00	7439-89-6		
Manganese	0.10	mg/L	0.010		1		10/02/18 12:00	7439-96-5		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: FE/MN 10/1
Pace Project No.: 7066406

QC Batch: 85211 Analysis Method: EPA 200.7
QC Batch Method: EPA 200.7 Analysis Description: 200.7 MET No Prep Drinking Water
Associated Lab Samples: 7066406001, 7066406002, 7066406003

METHOD BLANK: 391721 Matrix: Drinking Water
Associated Lab Samples: 7066406001, 7066406002, 7066406003

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Iron	mg/L	<0.020	0.020	10/02/18 11:50	
Manganese	mg/L	<0.010	0.010	10/02/18 11:50	

LABORATORY CONTROL SAMPLE: 391722

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Iron	mg/L	2	2.0	102	85-115	
Manganese	mg/L	.25	0.26	102	85-115	

MATRIX SPIKE SAMPLE: 391725

Parameter	Units	7066406001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Iron	mg/L	0.94	2	3.1	106	70-130	
Manganese	mg/L	0.15	.25	0.41	106	70-130	

MATRIX SPIKE SAMPLE: 391727

Parameter	Units	7066406002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Iron	mg/L	0.49	2	2.6	106	70-130	
Manganese	mg/L	0.061	.25	0.32	105	70-130	

SAMPLE DUPLICATE: 391724

Parameter	Units	7066406001 Result	Dup Result	RPD	Max RPD	Qualifiers
Iron	mg/L	0.94	0.94	0	20	
Manganese	mg/L	0.15	0.15	0	20	

SAMPLE DUPLICATE: 391726

Parameter	Units	7066406002 Result	Dup Result	RPD	Max RPD	Qualifiers
Iron	mg/L	0.49	0.47	2	20	
Manganese	mg/L	0.061	0.059	3	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: FE/MN 10/1

Pace Project No.: 7066406

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE


Project: FE/MN 10/1
Pace Project No.: 7066406

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
7066406001	S-108065	EPA 200.7	85211		
7066406002	S-108066	EPA 200.7	85211		
7066406003	COMB. ENTRY POINT	EPA 200.7	85211		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

WO#: 7066406



7066406

Sample Request Form PUBLIC WATER SUPPLIER

Date: 10-1-18

Collected By: W. B. ...

Accepted By: ...

Cooler Temp: 2.8 °C

WELL OFF LINE

WELL RUN TO SYSTEM

YES NO VOC'S PRESERVED WITH HCl

Client Info:

Name or Code: HAMPTON BAYS WATER DISTRICT
Address: P.O. BOX 1013
HAMPTON BAYS, NEW YORK 11946
(631) 728-0179


Phone #: _____
Attn: _____
Proj. # or (Name): _____
Bill To: _____
Copies To: _____

Sample Info:

Date/Time Collected:	Sample Type	Location	Origin	Treatment Type	Purpose	Field Readings Cl ₂ pH/Temp	Analysis	Lab No.
10:31 10-1-18	GW	WELL 4-1	RW	- S			IRON, MANG.	001
10:33 10-1-18	GW	WELL 4-2	RW	- S			IRON, MANG.	002
10:35 10-1-18	PW	Coma. Entry Point	D	- S	102	7.23 15.90C	IRON, MANG.	003

Sample Types	Purpose	Origin	Treatment Types
PW - Potable Water	RO - Routine	D - Distribution	AST - Air Stripper
GW - Groundwater	RE - Resample	RW - Raw Well	GAC - Granular Activated Charcoal
SW - Surface Water	S - Special	TW - Treated Well	N - Nitrate Removal Plant
WW - Waste Water		T - Tank	FE - Iron Removal Plant
AQ - Aqueous		MW - Monitoring Well	O - Other
S - Soil		I - Influent	
		E - Effluent	

Remarks: STU- REMEMBER THAT I NEED A RUSH ON THESE
NEED BY WED. AFTERNOON. THX - WATTEN



Sample Condition Upon Receipt

WO#: 7066406
PM: SWM Due Date: 10/09/18
CLIENT: HBW

Client Name: HBW

Courier: Fed Ex UPS USPS Client Commercial Pace Other

Tracking #:
Custody Seal on Cooler/Box Present: Yes No
Seals intact: Yes No

Packing Material: Bubble Wrap Bubble Bags Ziploc None Other
Temperature Blank Present: Yes No
Type of Ice: Wet Blue None

Thermometer Used: TH091 Correction Factor: 0.0
Cooler Temperature (°C): 2.8 Cooler Temperature Corrected (°C): 2.8
Date/Time 5035A kits placed in freezer

Temp should be above freezing to 6.0°C
USDA Regulated Soil (N/A, water sample)
Date and Initials of person examining contents: [Signature]

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? YES NO
Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (F-LI-C-010) and include with SCUR/COC paperwork.

Table with 16 rows and 3 columns: Question, Yes/No/N/A, and Comments. Includes items like Chain of Custody Present, Short Hold Time Analysis, Containers Intact, etc.

Client Notification/ Resolution:
Person Contacted:
Comments/ Resolution:

Field Data Required? Y / N
Date/Time:

* PM (Project Manager) review is documented electronically in LIMS.