

August 21, 2018

Rob King  
Hampton Bays Water District  
P.O. Box 1013  
Hampton Bays, NY 11946

RE: Project: DBP/CHLORATE 8/1  
Pace Project No.: 7060113

Dear Rob King:

Enclosed are the analytical results for sample(s) received by the laboratory on August 01, 2018. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Kimberley M. Mack for  
Stu Murrell  
stu.murrell@pacelabs.com  
(631)694-3040  
Project Manager

Enclosures

cc: Warren Booth, Hampton Bays Water District  
John Collins, H2M Group  
Stella Michaels, Hampton Bays Water District  
Paul Ponturo, H2M Group



## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## CERTIFICATIONS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

---

### Long Island Certification IDs

575 Broad Hollow Rd, Melville, NY 11747

New York Certification #: 10478 Primary Accrediting Body

New Jersey Certification #: NY158

Pennsylvania Certification #: 68-00350

Connecticut Certification #: PH-0435

Maryland Certification #: 208

Rhode Island Certification #: LAO00340

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

---

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## SAMPLE SUMMARY

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

---

Lab ID	Sample ID	Matrix	Date Collected	Date Received
7060113001	HB3	Drinking Water	08/01/18 08:05	08/01/18 16:45
7060113002	HB8	Drinking Water	08/01/18 09:45	08/01/18 16:45
7060113003	HB3	Drinking Water	08/01/18 08:05	08/01/18 16:45
7060113004	HB8	Drinking Water	08/01/18 09:45	08/01/18 16:45

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### SAMPLE ANALYTE COUNT

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

Lab ID	Sample ID	Method	Analysts	Analytes Reported
7060113001	HB3	EPA 552.2	KB1	7
		EPA 524.2	KGG	7
7060113002	HB8	EPA 552.2	KB1	7
		EPA 524.2	KGG	7
7060113003	HB3	EPA 300.1	BNK	2
7060113004	HB8	EPA 300.1	BNK	2

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

**Sample: HB3**      **Lab ID: 7060113001**      Collected: 08/01/18 08:05      Received: 08/01/18 16:45      Matrix: Drinking Water

Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>Field Chlorine and pH</b>									
Analytical Method:									
Field Chlorine, Free	<b>0.63</b>	mg/L			1		08/01/18 08:05		N3
<b>552.2 Haloacetic Acids</b>									
Analytical Method: EPA 552.2      Preparation Method: EPA 552.2									
Dibromoacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 22:39	631-64-1	
Dichloroacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 22:39	79-43-6	
Haloacetic Acids (Total)	<2.0	ug/L	2.0		1	08/07/18 10:22	08/17/18 22:39		
Monobromoacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 22:39	79-08-3	
Monochloroacetic Acid	<2.0	ug/L	2.0		1	08/07/18 10:22	08/17/18 22:39	79-11-8	
Trichloroacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 22:39	76-03-9	
<b>Surrogates</b>									
2,3-Dibromopropanoic Acid (S)	100	%	70-130		1	08/07/18 10:22	08/17/18 22:39	600-05-5	
<b>524.2 MSV</b>									
Analytical Method: EPA 524.2									
Bromodichloromethane	<b>0.66</b>	ug/L	0.50		1		08/03/18 10:17	75-27-4	
Bromoform	<0.50	ug/L	0.50		1		08/03/18 10:17	75-25-2	
Chloroform	<b>1.5</b>	ug/L	0.50		1		08/03/18 10:17	67-66-3	
Dibromochloromethane	<b>0.57</b>	ug/L	0.50		1		08/03/18 10:17	124-48-1	
Total Trihalomethanes (Calc.)	<b>2.7</b>	ug/L	0.50		1		08/03/18 10:17		
<b>Surrogates</b>									
1,2-Dichlorobenzene-d4 (S)	91	%	70-130		1		08/03/18 10:17	2199-69-1	
4-Bromofluorobenzene (S)	91	%	70-130		1		08/03/18 10:17	460-00-4	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### ANALYTICAL RESULTS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

**Sample: HB8**      **Lab ID: 7060113002**      Collected: 08/01/18 09:45      Received: 08/01/18 16:45      Matrix: Drinking Water

Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>Field Chlorine and pH</b>									
Analytical Method:									
Field Chlorine, Free	<b>0.55</b>	mg/L			1		08/01/18 09:45		N3
<b>552.2 Haloacetic Acids</b>									
Analytical Method: EPA 552.2      Preparation Method: EPA 552.2									
Dibromoacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 23:04	631-64-1	
Dichloroacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 23:04	79-43-6	
Haloacetic Acids (Total)	<2.0	ug/L	2.0		1	08/07/18 10:22	08/17/18 23:04		
Monobromoacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 23:04	79-08-3	
Monochloroacetic Acid	<2.0	ug/L	2.0		1	08/07/18 10:22	08/17/18 23:04	79-11-8	
Trichloroacetic Acid	<1.0	ug/L	1.0		1	08/07/18 10:22	08/17/18 23:04	76-03-9	
<b>Surrogates</b>									
2,3-Dibromopropanoic Acid (S)	86	%	70-130		1	08/07/18 10:22	08/17/18 23:04	600-05-5	
<b>524.2 MSV</b>									
Analytical Method: EPA 524.2									
Bromodichloromethane	<0.50	ug/L	0.50		1		08/03/18 10:45	75-27-4	
Bromoform	<0.50	ug/L	0.50		1		08/03/18 10:45	75-25-2	
Chloroform	1.2	ug/L	0.50		1		08/03/18 10:45	67-66-3	
Dibromochloromethane	<0.50	ug/L	0.50		1		08/03/18 10:45	124-48-1	
Total Trihalomethanes (Calc.)	1.2	ug/L	0.50		1		08/03/18 10:45		
<b>Surrogates</b>									
1,2-Dichlorobenzene-d4 (S)	91	%	70-130		1		08/03/18 10:45	2199-69-1	
4-Bromofluorobenzene (S)	92	%	70-130		1		08/03/18 10:45	460-00-4	

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

Sample: HB3		Lab ID: 7060113003		Collected: 08/01/18 08:05	Received: 08/01/18 16:45	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual
<b>300.1 Oxihalide IC Anions 28d</b>		Analytical Method: EPA 300.1							
Chlorate	<b>87.8</b>	ug/L	20.0		1		08/06/18 00:00	7790-93-4	
<b>Surrogates</b>									
Dichloroacetate (S)	99	%	90-115		1		08/06/18 00:00	79-43-6	

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## ANALYTICAL RESULTS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

Sample: HB8		Lab ID: 7060113004		Collected: 08/01/18 09:45	Received: 08/01/18 16:45	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual	
<b>300.1 Oxihalide IC Anions 28d</b>		Analytical Method: EPA 300.1								
Chlorate	<b>52.5</b>	ug/L	20.0		1		08/06/18 00:00	7790-93-4		
<b>Surrogates</b>										
Dichloroacetate (S)	93	%	90-115		1		08/06/18 00:00	79-43-6		

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.



**QUALITY CONTROL DATA**

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

QC Batch: 77676 Analysis Method: EPA 524.2  
 QC Batch Method: EPA 524.2 Analysis Description: 524.2 MSV  
 Associated Lab Samples: 7060113001, 7060113002

METHOD BLANK: 357047 Matrix: Water

Associated Lab Samples: 7060113001, 7060113002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Bromodichloromethane	ug/L	<0.50	0.50	08/03/18 07:01	
Bromoform	ug/L	<0.50	0.50	08/03/18 07:01	
Chloroform	ug/L	<0.50	0.50	08/03/18 07:01	
Dibromochloromethane	ug/L	<0.50	0.50	08/03/18 07:01	
Total Trihalomethanes (Calc.)	ug/L	<0.50	0.50	08/03/18 07:01	
1,2-Dichlorobenzene-d4 (S)	%	91	70-130	08/03/18 07:01	
4-Bromofluorobenzene (S)	%	91	70-130	08/03/18 07:01	

LABORATORY CONTROL SAMPLE: 357048

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Bromodichloromethane	ug/L	10	9.0	90	70-130	
Bromoform	ug/L	10	8.6	86	70-130	
Chloroform	ug/L	10	8.9	89	70-130	
Dibromochloromethane	ug/L	10	9.2	92	70-130	
Total Trihalomethanes (Calc.)	ug/L		35.7			
1,2-Dichlorobenzene-d4 (S)	%			98	70-130	
4-Bromofluorobenzene (S)	%			93	70-130	

SAMPLE DUPLICATE: 359963

Parameter	Units	7060360001 Result	Dup Result	RPD	Max RPD	Qualifiers
Bromodichloromethane	ug/L	<0.50	<0.50			20
Bromoform	ug/L	<0.50	<0.50			20
Chloroform	ug/L	<0.50	<0.50			20
Dibromochloromethane	ug/L	<0.50	<0.50			20
Total Trihalomethanes (Calc.)	ug/L	<0.50	<0.50			20
1,2-Dichlorobenzene-d4 (S)	%	84	87	4		20
4-Bromofluorobenzene (S)	%	86	89	4		20

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

**REPORT OF LABORATORY ANALYSIS**

This report shall not be reproduced, except in full,  
 without the written consent of Pace Analytical Services, LLC.

### QUALITY CONTROL DATA

Project: DBP/CHLORATE 8/1  
Pace Project No.: 7060113

QC Batch: 78069 Analysis Method: EPA 552.2  
QC Batch Method: EPA 552.2 Analysis Description: 5522 Haloacetic Acids  
Associated Lab Samples: 7060113001, 7060113002

METHOD BLANK: 358851 Matrix: Water  
Associated Lab Samples: 7060113001, 7060113002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Dibromoacetic Acid	ug/L	<1.0	1.0	08/17/18 21:22	
Dichloroacetic Acid	ug/L	<1.0	1.0	08/17/18 21:22	
Haloacetic Acids (Total)	ug/L	<2.0	2.0	08/17/18 21:22	
Monobromoacetic Acid	ug/L	<1.0	1.0	08/17/18 21:22	
Monochloroacetic Acid	ug/L	<2.0	2.0	08/17/18 21:22	
Trichloroacetic Acid	ug/L	<1.0	1.0	08/17/18 21:22	
2,3-Dibromopropanoic Acid (S)	%	95	70-130	08/17/18 21:22	

LABORATORY CONTROL SAMPLE: 358852

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dibromoacetic Acid	ug/L	6.2	4.6	73	70-130	
Dichloroacetic Acid	ug/L	6.2	4.5	72	70-130	
Haloacetic Acids (Total)	ug/L		27.3			
Monobromoacetic Acid	ug/L	6.2	4.4	71	70-130	
Monochloroacetic Acid	ug/L	12.5	9.3	74	70-130	
Trichloroacetic Acid	ug/L	6.2	4.5	72	70-130	
2,3-Dibromopropanoic Acid (S)	%			89	70-130	

LABORATORY CONTROL SAMPLE: 358853

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dibromoacetic Acid	ug/L	1	<1.0	81	70-130	
Dichloroacetic Acid	ug/L	1	<1.0	88	70-130	
Haloacetic Acids (Total)	ug/L		<2.0			
Monobromoacetic Acid	ug/L	1	<1.0	100	70-130	
Monochloroacetic Acid	ug/L	2	<2.0	82	70-130	
Trichloroacetic Acid	ug/L	1	<1.0	80	70-130	
2,3-Dibromopropanoic Acid (S)	%			90	70-130	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 358945 358946

Parameter	Units	7060223001		MSD		MSD		% Rec		Limits	RPD	Max RPD	Qual
		MS Result	Spike Conc.	Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec					
Dibromoacetic Acid	ug/L	<1.0	6.2	6.2	4.9	5.0	74	75	70-130	1	20		
Dichloroacetic Acid	ug/L	<1.0	6.2	6.2	4.8	4.8	77	77	70-130	0	20		
Haloacetic Acids (Total)	ug/L	<2.0			29.0	29.5				2	20		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### QUALITY CONTROL DATA

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

Parameter	Units	358945		358946		MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
		7060223001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result							
Monobromoacetic Acid	ug/L	<1.0	6.2	6.2	4.8	4.8	78	77	70-130	0	20	
Monochloroacetic Acid	ug/L	<2.0	12.5	12.5	9.8	10.2	79	81	70-130	4	20	
Trichloroacetic Acid	ug/L	<1.0	6.2	6.2	4.6	4.7	72	74	70-130	3	20	
2,3-Dibromopropanoic Acid (S)	%						94	93	70-130			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### QUALITY CONTROL DATA

Project: DBP/CHLORATE 8/1  
Pace Project No.: 7060113

QC Batch: 77984 Analysis Method: EPA 300.1  
QC Batch Method: EPA 300.1 Analysis Description: 300.1 Oxihalides IC Anions  
Associated Lab Samples: 7060113003, 7060113004

METHOD BLANK: 358540 Matrix: Water  
Associated Lab Samples: 7060113003, 7060113004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chlorate	ug/L	<20.0	20.0	08/06/18 00:00	
Dichloroacetate (S)	%	95	90-115	08/06/18 00:00	

LABORATORY CONTROL SAMPLE: 358541

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chlorate	ug/L	20	20.6	103	85-115	
Dichloroacetate (S)	%			97	90-115	

MATRIX SPIKE SAMPLE: 358543

Parameter	Units	7059389001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Chlorate	ug/L	133	60	187	90	75-115	
Dichloroacetate (S)	%				102	90-115	

SAMPLE DUPLICATE: 358542

Parameter	Units	7059389001 Result	Dup Result	RPD	Max RPD	Qualifiers
Chlorate	ug/L	133	131	2	10	
Dichloroacetate (S)	%	106	102	4		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

## QUALIFIERS

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

---

### DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

### ANALYTE QUALIFIERS

N3 Accreditation is not offered by the relevant laboratory accrediting body for this parameter.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

### QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: DBP/CHLORATE 8/1

Pace Project No.: 7060113

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
7060113001	HB3		77594		
7060113002	HB8		77594		
7060113001	HB3	EPA 552.2	78069	EPA 552.2	78352
7060113002	HB8	EPA 552.2	78069	EPA 552.2	78352
7060113001	HB3	EPA 524.2	77676		
7060113002	HB8	EPA 524.2	77676		
7060113003	HB3	EPA 300.1	77984		
7060113004	HB8	EPA 300.1	77984		

### REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.

WO#: 7060113



11747  
36

**Client Info:**

Name or Code: HAMPTON BAYS WATER DISTRICT  
Address: P.O. BOX 1013  
HAMPTON BAYS, NEW YORK 11946  
(631) 728-0179

Phone #: \_\_\_\_\_  
Attn: \_\_\_\_\_  
Proj. # or (Name): \_\_\_\_\_  
Bill To: \_\_\_\_\_  
Copies To: \_\_\_\_\_

**Sample Info:**

Date/Time Collected:	Sample Type	Location	Origin	Treatment Type	Purpose	Field Readings Cl <sub>2</sub>	pH/Temp	Analysis	Lab No.
8:05 AM 8-1-18	PW	# 3	D	-	RO	.63	7.34	TAM   HADS	001
9:45 AM 8-1-18	PW	# 8	D	-	RO	.55	7.57	TAM   HADS	002
8:05 AM 8-1-18	PW	# 3	D	-	RO	.63	7.34	CHLORATE	003
9:45 AM 8-1-18	PW	# 8	D	-	RO	.55	7.57	CHLORATE	004

**Sample Request Form  
PUBLIC WATER SUPPLIER**

Date: 8-1-18

Collected By: K. Tyndall

Accepted By: [Signature]

Cooler Temp: 3.7 °C

WELL OFF LINE

WELL RUN TO SYSTEM

YES  NO VOC'S PRESERVED WITH HCl

Sample Types	Purpose	Origin	Treatment Types
PW - Potable Water	RO - Routine	D - Distribution	AST - Air Stripper
GW - Groundwater	RE - Resample	RW - Raw Well	GAC - Granular Activated Charcoal
SW - Surface Water	S - Special	TW - Treated Well	N - Nitrate Removal Plant
WW - Waste Water		T - Tank	FE - Iron Removal Plant
AQ - Aqueous		MW - Monitoring Well	O - Other
S - Soil		I - Influent	
		E - Effluent	



**WO#: 7060113**  
 PM: SWM Due Date: 08/10/18  
 CLIENT: HBW

Client Name: HBW

Courier:  Fed Ex  UPS  USPS  Client  Commercial  Pace  Other

Tracking #: \_\_\_\_\_  
 Custody Seal on Cooler/Box Present:  Yes  No      Seals intact:  Yes  No

Packing Material:  Bubble Wrap  Bubble Bags  Ziploc  None  Other

Thermometer Used: TH091      Correction Factor: 0.0  
 Cooler Temperature (°C): 3.7      Cooler Temperature Corrected (°C): 3.7

Temp should be above freezing to 6.0°C  
 USDA Regulated Soil (  N/A, water sample)

Temperature Blank Present:  Yes  No

Type of Ice: Wet Blue None

Samples on ice, cooling process has begun  
 Date/Time 5035A kits placed in freezer \_\_\_\_\_

Date and Initials of person examining contents: Bohlio

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)?  YES  NO

Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)?  Yes  No

If Yes to either question, fill out a Regulated Soil Checklist (F-LI-C-010) and include with SCUR/COC paperwork.

			COMMENTS:
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	1.
Chain of Custody Filled Out:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	2.
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	3.
Sampler Name & Signature on COC:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested:	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD)	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	8.
Correct Containers Used:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	9.
-Pace Containers Used:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	10.
Containers Intact:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	11. Note if sediment is visible in the dissolved container.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes	<input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	12.
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No	13. <input type="checkbox"/> HNO <sub>3</sub> <input type="checkbox"/> H <sub>2</sub> SO <sub>4</sub> <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
-Includes date/time/ID/Analysis Matrix <u>ST</u> <u>WT</u> <u>OIL</u>			Sample #
All containers needing preservation have been checked	<input type="checkbox"/> Yes	<input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Initial when completed: _____ Lot # of added preservative: _____ Date/Time preservative added: _____
pH paper Lot #			
All containers needing preservation are found to be in compliance with EPA recommendation? (HNO <sub>3</sub> , H <sub>2</sub> SO <sub>4</sub> , HCl, NaOH > 9 Sulfide, NaOH > 12 Cyanide)	<input type="checkbox"/> Yes	<input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis			
Samples checked for dechlorination:	<input type="checkbox"/> Yes	<input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot #			
Residual chlorine strips Lot #			
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	15.
Trip Blank Present:	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	16.
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes	<input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Pace Trip Blank Lot # (if applicable): _____			

Field Data Required? Y / N  
 Date/Time: \_\_\_\_\_

Client Notification/ Resolution: \_\_\_\_\_  
 Person Contacted: \_\_\_\_\_  
 Comments/ Resolution: \_\_\_\_\_